

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2072

Name: RITTER 1  
Sterilizer ID: V2178198  
06/29/2022 08:59 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 154.7 F 0.0  
2:00 157.6 F 1.2  
4:00 173.9 F 4.2  
6:00 192.5 F 8.5  
8:00 211.0 F 13.4  
10:00 218.3 F 6.2  
12:00 231.8 F 11.0  
14:00 231.1 F 7.3  
16:00 244.1 F 13.3  
18:00 242.7 F 11.7  
20:00 255.3 F 18.1  
22:00 266.5 F 24.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.4  
1:00 272.6 F 29.2  
2:00 272.4 F 28.8  
3:00 271.6 F 28.3  
4:00 272.1 F 29.0

Min 271.5 F 28.3  
Max 272.7 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 182.3 F 0.0  
5:00 207.9 F 0.0  
10:00 223.3 F 0.0  
15:00 233.3 F 0.0  
20:00 233.0 F 0.0  
25:00 228.8 F 0.0  
30:00 218.2 F 0.0  
Drying Complete

FILLING: 0:49  
HEATING: 23:42  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 01:01:40  
CYCLE COMPLETE

06/29/2022 10:00 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_