
Midmark M11 - v1.0.5
Total Cycles: 4062

Name: RITTER2
Sterilizer ID: V2313228
10/29/2024 10:20 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 158.5 F 0.0
2:00 158.6 F 1.6
4:00 172.1 F 5.0
6:00 196.4 F 9.1
8:00 212.2 F 13.4
10:00 220.7 F 7.3
12:00 232.4 F 11.9
14:00 233.3 F 8.9
16:00 241.0 F 10.8
18:00 245.7 F 13.5
20:00 257.3 F 19.8
22:00 267.7 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.0
1:00 272.4 F 29.5
2:00 271.2 F 29.1
3:00 272.2 F 29.5
4:00 271.1 F 28.9

Min 271.1 F 28.8
Max 272.6 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.3 F 0.0
5:00 204.1 F 0.0
10:00 220.5 F 0.0
15:00 221.3 F 0.0
20:00 221.3 F 0.0
25:00 220.1 F 0.0
30:00 217.3 F 0.0
Drying Complete

FILLING: 1:21
HEATING: 23:22
STERILIZING: 4:00
VENTING: 3:47
DRYING: 30:00
TOTAL CYCLE: 01:02:30
CYCLE COMPLETE

10/29/2024 11:22 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____