

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1853

Name: RITTER2  
Sterilizer ID: V2313228  
12/19/2022 10:44 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 203.3 F 0.0  
2:00 192.8 F 2.8  
4:00 197.7 F 8.3  
6:00 217.6 F 9.5  
8:00 230.6 F 10.2  
10:00 233.8 F 7.9  
12:00 240.7 F 10.8  
14:00 246.6 F 14.3  
16:00 259.7 F 21.6  
18:00 271.4 F 29.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 29.4  
1:00 272.8 F 30.0  
2:00 271.7 F 29.3  
3:00 272.8 F 30.2  
4:00 272.2 F 29.6  
  
Min 271.6 F 29.2  
Max 272.8 F 30.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.6 F 0.0  
5:00 227.1 F 0.0  
10:00 234.9 F 0.0  
15:00 229.5 F 0.0  
20:00 225.5 F 0.0  
25:00 224.7 F 0.0  
30:00 223.6 F 0.0  
Drying Complete

FILLING: 2:31  
HEATING: 18:40  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 00:58:17  
CYCLE COMPLETE

12/19/2022 11:42 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_