
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 004302
12:50 06/09/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:39
PRESSURIZING 2:45
STERILIZING 3:46
136.5 C 222kPa 3:46
136.4 C 225kPa 4:17
136.4 C 223kPa 4:47
CYCLE ABORTED 5:07
STOP BUTTON PRESSED
NOT STERILE
END TIME 5:07

Digital Signature #
5071541193EA8072
