

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1836

Name: RITTER2  
Sterilizer ID: V2313228  
12/14/2022 08:44 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 78.5 F 0.0  
2:00 80.8 F 0.0  
4:00 90.9 F 0.0  
6:00 127.9 F 0.0  
8:00 151.1 F 0.0  
10:00 184.6 F 0.6  
12:00 202.5 F 4.2  
14:00 215.2 F 7.6  
16:00 226.9 F 11.4  
18:00 228.3 F 6.3  
20:00 239.5 F 11.4  
22:00 236.5 F 9.1  
24:00 240.3 F 10.1  
26:00 247.9 F 14.2  
28:00 259.6 F 20.7  
30:00 270.2 F 27.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.5  
1:00 272.5 F 29.1  
2:00 271.5 F 28.6  
3:00 272.5 F 29.1  
4:00 271.5 F 28.4

Min 271.5 F 28.3  
Max 272.6 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.1 F 0.0  
5:00 206.4 F 0.0  
10:00 218.3 F 0.0  
15:00 225.6 F 0.0  
20:00 223.5 F 0.0  
25:00 225.3 F 0.0  
30:00 226.5 F 0.0  
Drying Complete

FILLING: 0:48  
HEATING: 30:55  
STERILIZING: 4:00  
VENTING: 3:03  
DRYING: 30:00  
TOTAL CYCLE: 01:08:47  
CYCLE COMPLETE

12/14/2022 09:52 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_