
Midmark M11 - v1.0.5
Total Cycles: 2860

Name: RITTER2
Sterilizer ID: V2313228
09/11/2023 12:00 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 171.7 F 0.0
2:00 172.2 F 3.1
4:00 187.8 F 7.5
6:00 207.3 F 12.3
8:00 225.5 F 7.4
10:00 234.9 F 12.4
12:00 237.0 F 9.6
14:00 237.4 F 9.8
16:00 247.2 F 14.0
18:00 258.2 F 20.0
20:00 268.2 F 26.4
22:00 271.7 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.3
1:00 271.8 F 29.2
2:00 272.2 F 29.3
3:00 272.4 F 29.5
4:00 272.6 F 29.6

Min 271.6 F 29.2
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.1 F 0.0
5:00 233.9 F 0.0
10:00 215.1 F 0.0
15:00 211.2 F 0.0
20:00 208.2 F 0.0
25:00 207.5 F 0.0
30:00 195.7 F 0.0
Drying Complete

FILLING: 1:42
HEATING: 22:02
STERILIZING: 4:00
VENTING: 3:43
DRYING: 30:00
TOTAL CYCLE: 01:01:27
CYCLE COMPLETE

09/11/2023 01:01 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____