
Midmark M11 - v1.0.5
Total Cycles: 3686

Name: RITTER2
Sterilizer ID: V2313228
06/11/2024 02:23 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 84.3 F 0.0
2:00 87.0 F 0.2
4:00 108.5 F 1.5
6:00 134.5 F 4.0
8:00 161.5 F 7.0
10:00 182.1 F 10.1
12:00 199.7 F 8.5
14:00 212.2 F 5.4
16:00 224.9 F 8.8
18:00 235.4 F 12.9
20:00 233.8 F 9.2
22:00 241.8 F 11.1
24:00 244.1 F 12.8
26:00 255.1 F 18.7
28:00 265.3 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.1
1:00 273.0 F 29.8
2:00 271.9 F 29.0
3:00 272.3 F 29.7
4:00 272.8 F 29.6

Min 271.6 F 28.9
Max 273.0 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.6 F 0.0
5:00 193.3 F 0.0
10:00 199.3 F 0.0
15:00 190.7 F 0.0
20:00 197.0 F 0.0
25:00 201.5 F 0.0
30:00 203.3 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 29:57
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:08:25
CYCLE COMPLETE

06/11/2024 03:31 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____