

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1463

Name: RITTER 1  
Sterilizer ID: V2178198  
12/15/2021 06:09 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 131.3 F 0.0  
2:00 143.5 F 0.0  
4:00 154.3 F 1.3  
6:00 177.2 F 3.5  
8:00 199.6 F 7.1  
10:00 213.4 F 11.2  
12:00 224.2 F 5.3  
14:00 234.3 F 10.2  
16:00 236.5 F 8.4  
18:00 243.6 F 12.6  
20:00 241.8 F 11.0  
22:00 254.6 F 17.5  
24:00 265.8 F 24.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.0  
1:00 272.5 F 28.8  
2:00 271.9 F 28.2  
3:00 271.2 F 28.1  
4:00 272.5 F 28.9

Min 271.2 F 27.9  
Max 272.5 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.2 F 0.0  
5:00 218.2 F 0.0  
10:00 219.9 F 0.0  
15:00 218.4 F 0.0  
20:00 216.7 F 0.0  
25:00 215.8 F 0.0  
30:00 215.2 F 0.0  
Drying Complete

FILLING: 0:52  
HEATING: 25:44  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 01:03:51  
CYCLE COMPLETE

12/16/2021 07:58 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02