
Midmark M11 - v1.0.5
Total Cycles: 4389

Name: RITTER2
Sterilizer ID: V2313228
02/28/2025 10:14 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.6 F 0.0
2:00 69.8 F 0.2
4:00 73.2 F 0.9
6:00 88.0 F 2.7
8:00 118.9 F 5.4
10:00 159.1 F 8.3
12:00 167.8 F 11.8
14:00 203.6 F 3.1
16:00 215.6 F 6.4
18:00 225.4 F 9.7
20:00 232.4 F 13.4
22:00 232.9 F 9.2
24:00 241.9 F 12.2
26:00 241.9 F 11.9
28:00 252.6 F 17.4
30:00 262.5 F 23.1
32:00 271.5 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 30.0
1:00 272.4 F 30.1
2:00 271.9 F 30.1
3:00 272.2 F 30.1
4:00 271.8 F 29.9

Min 271.7 F 29.8
Max 272.5 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.9 F 0.0
5:00 211.3 F 0.0
10:00 209.2 F 0.0
15:00 196.6 F 0.0
20:00 183.7 F 0.0
25:00 178.2 F 0.0
30:00 177.4 F 0.0
Drying Complete

FILLING: 0:36
HEATING: 33:14
STERILIZING: 4:00
VENTING: 3:48
DRYING: 30:00
TOTAL CYCLE: 01:11:38
CYCLE COMPLETE

02/28/2025 11:25 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____