

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3909

Name: RITTER 1  
Sterilizer ID: V2178198  
10/17/2023 08:03 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.8 F 0.0  
2:00 71.5 F 0.0  
4:00 81.6 F 0.0  
6:00 120.6 F 0.1  
8:00 152.6 F 1.0  
10:00 174.1 F 4.3  
12:00 193.1 F 8.0  
14:00 213.8 F 11.6  
16:00 220.7 F 4.5  
18:00 229.7 F 8.6  
20:00 239.4 F 12.9  
22:00 237.2 F 9.6  
24:00 237.9 F 9.3  
26:00 247.2 F 13.7  
28:00 258.9 F 20.2  
30:00 269.7 F 27.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.2  
1:00 272.4 F 28.7  
2:00 271.5 F 28.5  
3:00 271.9 F 28.5  
4:00 271.8 F 28.7

Min 271.1 F 28.1  
Max 272.6 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 173.8 F 0.0  
5:00 231.8 F 0.0  
10:00 232.4 F 0.0  
15:00 233.6 F 0.0  
20:00 228.3 F 0.0  
25:00 230.1 F 0.0  
30:00 192.7 F 0.0  
Drying Complete

FILLING: 0:50  
HEATING: 30:56  
STERILIZING: 4:00  
VENTING: 3:21  
DRYING: 30:00  
TOTAL CYCLE: 01:09:07  
CYCLE COMPLETE

10/17/2023 09:12 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_