
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 007135
11:52 28/03/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:03
PRESSURIZING 2:06
STERILIZING 3:11
136.4 C 223kPa 3:11
136.4 C 225kPa 3:41
136.4 C 224kPa 4:11
136.3 C 222kPa 4:41
136.4 C 222kPa 5:11
136.4 C 222kPa 5:41
136.1 C 219kPa 6:11
136.4 C 224kPa 6:41
136.4 C 224kPa 6:41
VENTING 6:41
AIR DRYING 7:32
CYCLE COMPLETE 7:41

Digital Signature #
D443700B0902B98F
