
Midmark M11 - v1.0.0
Total Cycles: 4066

Name: RITTER 1
Sterilizer ID: V2178198
11/29/2023 01:30 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

mm:ss	Degrees	PSI
0:00	141.8 F	0.0
2:00	144.3 F	2.0
4:00	163.4 F	5.1
6:00	185.6 F	8.6
8:00	214.4 F	13.1
10:00	224.8 F	6.1
12:00	228.4 F	10.7
14:00	231.2 F	7.5
16:00	240.6 F	12.6
18:00	240.7 F	10.9
20:00	253.0 F	17.0
22:00	263.9 F	23.4
24:00	271.3 F	28.5

STERILIZING

mm:ss	Degrees	PSI
0:00	271.3 F	28.8
1:00	271.3 F	28.7
2:00	271.6 F	28.8
3:00	272.2 F	29.0
4:00	272.4 F	29.2

Min 271.2 F 28.7
Max 273.0 F 29.3

VENTING CHAMBER

DRYING

Drying Start	mm:ss	Degrees	PSI
	0:00	178.6 F	0.0
	5:00	203.8 F	0.0
	10:00	200.7 F	0.0
	15:00	197.4 F	0.0
	20:00	192.5 F	0.0
	25:00	189.2 F	0.0
Drying Complete	30:00	186.1 F	0.0

FILLING: 1:09
HEATING: 24:45
STERILIZING: 4:00
VENTING: 3:47
DRYING: 30:00
TOTAL CYCLE: 01:03:41
CYCLE COMPLETE

11/29/2023 02:33 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____