

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2754

Name: RITTER 1  
Sterilizer ID: V2178198  
01/09/2023 04:37 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 172.0 F 0.0  
2:00 172.8 F 0.0  
4:00 191.2 F 3.4  
6:00 209.2 F 7.9  
8:00 229.4 F 12.7  
10:00 232.0 F 8.5  
12:00 239.9 F 11.5  
14:00 242.5 F 11.9  
16:00 240.4 F 10.6  
18:00 253.9 F 17.3  
20:00 265.2 F 24.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.2  
1:00 272.4 F 28.9  
2:00 271.4 F 28.2  
3:00 272.1 F 29.0  
4:00 272.1 F 28.7  
  
Min 271.2 F 28.1  
Max 272.6 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 198.2 F 0.0  
5:00 234.7 F 0.0  
10:00 229.9 F 0.0  
15:00 230.3 F 0.0  
20:00 227.9 F 0.0  
25:00 226.3 F 0.0  
30:00 228.6 F 0.0  
Drying Complete

FILLING: 0:53  
HEATING: 21:51  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 00:59:60  
CYCLE COMPLETE

01/09/2023 05:36 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_