
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001683
11:41 11/01/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:18
PRESSURIZING 2:26
STERILIZING 3:27
136.0 C 220kPa 3:27
135.8 C 223kPa 3:57
136.4 C 225kPa 4:27
136.2 C 223kPa 4:57
136.4 C 226kPa 5:27
136.3 C 224kPa 5:57
136.2 C 222kPa 6:27
136.4 C 223kPa 6:57
136.4 C 223kPa 6:57
VENTING 6:57
AIR DRYING 7:48
CYCLE COMPLETE 8:06

Digital Signature #
CF4257B8A08D4EC2
