
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000269
8:44 01/10/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:05
PRESSURIZING 3:19
STERILIZING 4:54
136.3 C 225kPa 4:54
136.3 C 232kPa 5:24
136.3 C 231kPa 5:54
136.3 C 232kPa 6:24
136.4 C 233kPa 6:54
136.3 C 233kPa 7:24
136.3 C 232kPa 7:54
136.4 C 233kPa 8:24
136.4 C 233kPa 8:24
VENTING 8:24
AIR DRYING 9:25
CYCLE COMPLETE 24:25

Digital Signature #
77FEF1795EF10300
