

-----  
Midmark M11 - v1.0.0  
Total Cycles: 618

Name: MIDMARK  
Sterilizer ID: V2178198  
01/21/2021 12:57 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 130.3 F 0.0  
2:00 134.9 F 0.9  
4:00 166.4 F 3.2  
6:00 188.1 F 7.0  
8:00 206.2 F 11.9  
10:00 219.6 F 5.8  
12:00 232.7 F 10.9  
14:00 230.9 F 7.4  
16:00 243.9 F 13.5  
18:00 242.3 F 11.9  
20:00 255.2 F 18.5  
22:00 266.5 F 25.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 28.7  
1:00 272.0 F 28.9  
2:00 272.1 F 29.1  
3:00 272.4 F 29.3  
4:00 272.4 F 29.5

Min 271.3 F 28.6  
Max 272.6 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 165.6 F 0.0  
5:00 230.2 F 0.0  
10:00 243.3 F 0.0  
15:00 228.8 F 0.0  
20:00 220.2 F 0.0  
25:00 219.1 F 0.0  
30:00 218.3 F 0.0  
Drying Complete

FILLING: 1:03  
HEATING: 23:20  
STERILIZING: 4:00  
VENTING: 3:03  
DRYING: 30:00  
TOTAL CYCLE: 01:01:26  
CYCLE COMPLETE

01/21/2021 01:58 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_