
Midmark M11 - v1.0.0
Total Cycles: 3173

Name: RITTER 1
Sterilizer ID: V2178198
04/28/2023 07:44 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 77.6 F 0.0
2:00 82.9 F 0.0
4:00 116.5 F 0.0
6:00 136.8 F 0.0
8:00 157.7 F 1.6
10:00 178.9 F 4.5
12:00 194.6 F 7.9
14:00 206.9 F 11.2
16:00 215.0 F 5.1
18:00 226.8 F 8.1
20:00 237.5 F 12.3
22:00 235.4 F 8.7
24:00 242.4 F 11.1
26:00 245.2 F 12.6
28:00 256.9 F 18.7
30:00 267.5 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 27.9
1:00 272.2 F 28.2
2:00 272.5 F 28.7
3:00 271.5 F 27.8
4:00 272.6 F 28.5

Min 271.4 F 27.8
Max 272.7 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.5 F 0.0
5:00 197.0 F 0.0
10:00 209.4 F 0.0
15:00 212.1 F 0.0
20:00 213.0 F 0.0
25:00 212.4 F 0.0
30:00 212.5 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 31:22
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:09:22
CYCLE COMPLETE

04/28/2023 08:53 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____