
Midmark M11 - v1.0.0
Total Cycles: 4639

Name: RITTER 1
Sterilizer ID: V2178198
06/13/2024 04:50 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 159.9 F 0.0
2:00 159.3 F 3.3
4:00 176.7 F 6.7
6:00 194.8 F 10.5
8:00 212.6 F 5.7
10:00 223.1 F 7.6
12:00 234.1 F 11.7
14:00 232.2 F 7.8
16:00 244.4 F 13.0
18:00 241.9 F 11.3
20:00 254.2 F 17.4
22:00 264.8 F 23.8
24:00 271.4 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.4 F 29.1
2:00 272.1 F 28.5
3:00 271.3 F 28.5
4:00 272.4 F 29.2

Min 271.2 F 28.2
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.9 F 0.0
5:00 209.1 F 0.0
10:00 206.6 F 0.0
15:00 208.7 F 0.0
20:00 199.7 F 0.0
25:00 197.3 F 0.0
30:00 191.4 F 0.0
Drying Complete

FILLING: 1:08
HEATING: 24:01
STERILIZING: 4:00
VENTING: 3:49
DRYING: 30:00
TOTAL CYCLE: 01:02:58
CYCLE COMPLETE

06/13/2024 05:52 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____