
Midmark M11 - v1.0.5
Total Cycles: 147

Name: RITTER2
Sterilizer ID: V2313228
06/28/2021 09:47 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 186.9 F 0.0
2:00 184.0 F 3.3
4:00 198.7 F 8.8
6:00 219.3 F 8.6
8:00 230.2 F 10.1
10:00 234.1 F 8.6
12:00 241.5 F 11.3
14:00 247.2 F 14.2
16:00 259.5 F 21.1
18:00 270.4 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.0
1:00 272.3 F 29.2
2:00 272.3 F 29.8
3:00 271.7 F 28.9
4:00 272.7 F 29.6

Min 271.5 F 28.9
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.8 F 0.0
5:00 229.1 F 0.0
10:00 232.8 F 0.0
15:00 225.8 F 0.0
20:00 183.8 F 0.0
25:00 178.5 F 0.0
30:00 176.1 F 0.0
Drying Complete

FILLING: 2:22
HEATING: 18:47
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:58:18
CYCLE COMPLETE

06/28/2021 10:49 AM
ACCEPTED: X
REJECTED:
Operator: U01