

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 003835  
13:50 16/06/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:16  
PRESSURIZING 2:22  
STERILIZING 3:22  
136.0 C 220kPa 3:22  
135.9 C 222kPa 3:52  
136.4 C 224kPa 4:22  
136.3 C 224kPa 4:52  
136.4 C 225kPa 5:22  
136.5 C 225kPa 5:52  
136.4 C 224kPa 6:22  
136.2 C 225kPa 6:52  
136.2 C 225kPa 6:52  
VENTING 6:52  
AIR DRYING 7:40  
CYCLE COMPLETE 7:57

-----  
Digital Signature #  
D8B87E65DDF00D4A  
-----