
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001191
11:35 17/02/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:57
PRESSURIZING 2:00
STERILIZING 3:03
136.4 C 223kPa 3:03
136.2 C 223kPa 3:33
136.4 C 219kPa 4:03
136.4 C 223kPa 4:33
136.4 C 218kPa 5:03
136.4 C 224kPa 5:33
136.4 C 225kPa 6:03
136.1 C 221kPa 6:33
136.1 C 221kPa 6:33
VENTING 6:33
AIR DRYING 7:21
CYCLE COMPLETE 9:50

Digital Signature #
D1C740812B4417AE
