
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 005112
9:47 14/02/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:50
PRESSURIZING 2:54
STERILIZING 4:03
136.4 C 223kPa 4:03
136.1 C 219kPa 4:33
136.2 C 221kPa 5:03
136.4 C 223kPa 5:33
136.4 C 223kPa 6:03
136.2 C 221kPa 6:33
136.4 C 223kPa 7:03
136.0 C 220kPa 7:33
136.0 C 220kPa 7:33
VENTING 7:33
AIR DRYING 8:21
CYCLE COMPLETE 10:55

Digital Signature #
177FE25669E83B2E
