

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4712

Name: RITTER 1  
Sterilizer ID: V2178198  
07/10/2024 03:28 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 153.6 F 0.0  
2:00 153.2 F 2.6  
4:00 162.8 F 5.8  
6:00 194.1 F 9.3  
8:00 213.9 F 7.3  
10:00 220.0 F 7.1  
12:00 230.6 F 11.2  
14:00 232.5 F 7.0  
16:00 241.9 F 12.7  
18:00 240.5 F 11.0  
20:00 252.5 F 17.0  
22:00 263.3 F 23.2  
24:00 272.4 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.6  
1:00 272.3 F 29.3  
2:00 271.4 F 28.6  
3:00 271.7 F 29.1  
4:00 272.1 F 29.2

Min 271.0 F 28.5  
Max 272.5 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.3 F 0.0  
5:00 208.5 F 0.0  
10:00 212.6 F 0.0  
15:00 218.5 F 0.0  
20:00 226.9 F 0.0  
25:00 234.6 F 0.0  
30:00 233.1 F 0.0  
Drying Complete

FILLING: 1:02  
HEATING: 24:17  
STERILIZING: 4:00  
VENTING: 3:52  
DRYING: 30:00  
TOTAL CYCLE: 01:03:11  
CYCLE COMPLETE

07/10/2024 04:31 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_