
Midmark M11 - v1.0.5
Total Cycles: 2858

Name: RITTER2
Sterilizer ID: V2313228
09/11/2023 09:50 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.8 F 0.0
2:00 167.7 F 1.8
4:00 180.2 F 5.6
6:00 202.4 F 10.1
8:00 214.3 F 6.8
10:00 227.0 F 9.3
12:00 232.5 F 10.2
14:00 241.7 F 11.7
16:00 239.1 F 10.2
18:00 251.7 F 16.3
20:00 262.4 F 22.5
22:00 272.3 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.6
1:00 271.8 F 29.3
2:00 273.1 F 29.7
3:00 271.7 F 29.1
4:00 272.7 F 29.9

Min 270.9 F 28.4
Max 273.1 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.0 F 0.0
5:00 216.9 F 0.0
10:00 217.9 F 0.0
15:00 215.6 F 0.0
20:00 215.3 F 0.0
25:00 220.7 F 0.0
30:00 222.3 F 0.0
Drying Complete

FILLING: 1:41
HEATING: 22:22
STERILIZING: 4:00
VENTING: 3:33
DRYING: 30:00
TOTAL CYCLE: 01:01:36
CYCLE COMPLETE

09/11/2023 10:51 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____