
Midmark M11 - v1.0.5
Total Cycles: 1678

Name: RITTER2
Sterilizer ID: V2313228
11/03/2022 10:14 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 183.3 F 0.0
2:00 182.7 F 3.7
4:00 205.4 F 9.5
6:00 224.5 F 6.7
8:00 236.3 F 11.6
10:00 236.3 F 9.8
12:00 236.7 F 8.8
14:00 251.1 F 16.2
16:00 263.5 F 23.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.9
1:00 272.5 F 29.3
2:00 271.9 F 29.4
3:00 272.6 F 29.2
4:00 271.7 F 29.4

Min 271.3 F 28.8
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.7 F 0.0
5:00 237.4 F 0.0
10:00 207.5 F 0.0
15:00 197.3 F 0.0
20:00 214.7 F 0.0
25:00 225.6 F 0.0
30:00 215.4 F 0.0
Drying Complete

FILLING: 2:46
HEATING: 17:59
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:57:55
CYCLE COMPLETE

11/03/2022 11:11 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____