
Midmark M11 - v1.0.0
Total Cycles: 3424

Name: RITTER 1
Sterilizer ID: V2178198
06/21/2023 02:03 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 174.1 F 0.0
2:00 171.5 F 2.5
4:00 189.0 F 6.3
6:00 209.2 F 10.7
8:00 222.8 F 5.1
10:00 234.2 F 10.3
12:00 237.1 F 7.9
14:00 243.8 F 12.9
16:00 242.7 F 11.5
18:00 255.6 F 18.2
20:00 266.9 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.6 F 28.8
2:00 271.5 F 28.3
3:00 272.6 F 28.8
4:00 271.5 F 28.1

Min 271.5 F 28.1
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.3 F 0.0
5:00 234.1 F 0.0
10:00 223.2 F 0.0
15:00 221.2 F 0.0
20:00 220.1 F 0.0
25:00 220.8 F 0.0
30:00 219.8 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 21:32
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:59:42
CYCLE COMPLETE

06/21/2023 03:02 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____