

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3430

Name: RITTER 1  
Sterilizer ID: V2178198  
06/22/2023 11:50 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 175.1 F 0.0  
2:00 175.3 F 3.1  
4:00 188.5 F 7.0  
6:00 210.9 F 12.0  
8:00 220.3 F 6.5  
10:00 233.3 F 11.6  
12:00 233.7 F 8.5  
14:00 240.5 F 10.6  
16:00 245.9 F 13.2  
18:00 258.5 F 19.9  
20:00 269.4 F 26.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.3  
1:00 272.5 F 29.1  
2:00 271.9 F 28.5  
3:00 271.6 F 28.5  
4:00 272.6 F 29.1

Min 271.3 F 28.2  
Max 272.7 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.1 F 0.0  
5:00 214.5 F 0.0  
10:00 227.4 F 0.0  
15:00 231.4 F 0.0  
20:00 225.4 F 0.0  
25:00 222.5 F 0.0  
30:00 220.7 F 0.0  
Drying Complete

FILLING: 1:05  
HEATING: 21:03  
STERILIZING: 4:00  
VENTING: 3:19  
DRYING: 30:00  
TOTAL CYCLE: 00:59:26  
CYCLE COMPLETE

06/22/2023 12:49 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_