

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1277

Name: RITTER 1  
Sterilizer ID: V2178198  
10/19/2021 10:25 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 149.1 F 0.0  
2:00 148.8 F 1.6  
4:00 162.2 F 4.3  
6:00 191.0 F 8.1  
8:00 219.6 F 13.2  
10:00 227.2 F 7.8  
12:00 238.9 F 12.6  
14:00 235.5 F 9.7  
16:00 239.0 F 9.1  
18:00 246.9 F 14.0  
20:00 258.6 F 20.3  
22:00 269.3 F 27.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.6  
1:00 272.5 F 29.4  
2:00 271.8 F 28.7  
3:00 271.7 F 29.0  
4:00 272.5 F 29.3

Min 271.3 F 28.5  
Max 272.6 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.2 F 0.0  
5:00 218.0 F 0.0  
10:00 213.1 F 0.0  
15:00 211.7 F 0.0  
20:00 211.5 F 0.0  
25:00 209.5 F 0.0  
30:00 210.5 F 0.0  
Drying Complete

FILLING: 1:25  
HEATING: 23:06  
STERILIZING: 4:00  
VENTING: 3:21  
DRYING: 30:00  
TOTAL CYCLE: 01:01:53  
CYCLE COMPLETE

10/19/2021 11:27 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01