

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3516

Name: RITTER 1  
Sterilizer ID: V2178198  
07/12/2023 09:29 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 182.9 F 0.0  
2:00 178.3 F 0.7  
4:00 179.0 F 3.1  
6:00 191.2 F 6.6  
8:00 210.2 F 11.4  
10:00 220.9 F 5.6  
12:00 231.5 F 10.5  
14:00 232.4 F 7.8  
16:00 243.8 F 13.1  
18:00 242.7 F 11.7  
20:00 254.8 F 17.9  
22:00 265.3 F 24.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.4  
1:00 272.6 F 29.1  
2:00 271.7 F 28.5  
3:00 271.8 F 29.0  
4:00 272.1 F 29.0

Min 271.2 F 28.3  
Max 272.6 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.8 F 0.0  
5:00 207.1 F 0.0  
10:00 219.9 F 0.0  
15:00 225.7 F 0.0  
20:00 227.8 F 0.0  
25:00 228.4 F 0.0  
30:00 227.5 F 0.0  
Drying Complete

FILLING: 1:04  
HEATING: 23:54  
STERILIZING: 4:00  
VENTING: 3:30  
DRYING: 30:00  
TOTAL CYCLE: 01:02:28  
CYCLE COMPLETE

07/12/2023 10:31 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_