
Midmark M11 - v1.0.0
Total Cycles: 3529

Name: RITTER 1
Sterilizer ID: V2178198
07/14/2023 12:38 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 120.7 F 0.0
2:00 124.3 F 0.0
4:00 146.5 F 1.9
6:00 174.0 F 4.9
8:00 198.4 F 8.4
10:00 210.5 F 12.9
12:00 220.7 F 6.0
14:00 231.5 F 10.2
16:00 234.7 F 9.0
18:00 241.1 F 11.5
20:00 238.0 F 9.5
22:00 251.0 F 15.7
24:00 261.9 F 21.9
26:00 271.8 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.4 F 28.7
2:00 271.6 F 28.5
3:00 272.4 F 28.7
4:00 271.5 F 28.3

Min 271.4 F 28.1
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.9 F 0.0
5:00 204.2 F 0.0
10:00 212.0 F 0.0
15:00 214.7 F 0.0
20:00 212.2 F 0.0
25:00 211.1 F 0.0
30:00 209.2 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 26:33
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 01:04:46
CYCLE COMPLETE

07/14/2023 01:42 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____