
Midmark M11 - v1.0.0
Total Cycles: 2501

Name: RITTER 1
Sterilizer ID: V2178198
11/02/2022 11:49 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 192.3 F 0.0
2:00 190.5 F 4.5
4:00 202.1 F 9.1
6:00 214.2 F 8.5
8:00 225.9 F 8.1
10:00 238.2 F 13.2
12:00 237.7 F 10.2
14:00 237.6 F 8.1
16:00 249.0 F 14.7
18:00 260.8 F 21.3
20:00 271.5 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.4 F 28.8
2:00 271.8 F 28.7
3:00 272.3 F 28.7
4:00 271.7 F 28.6

Min 271.5 F 28.2
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.1 F 0.0
5:00 237.3 F 0.0
10:00 233.6 F 0.0
15:00 227.1 F 0.0
20:00 219.7 F 0.0
25:00 225.1 F 0.0
30:00 227.0 F 0.0
Drying Complete

FILLING: 1:10
HEATING: 20:36
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:58:56
CYCLE COMPLETE

11/02/2022 12:47 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____