
Midmark M11 - v1.0.0
Total Cycles: 2787

Name: RITTER 1
Sterilizer ID: V2178198
01/17/2023 03:17 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 195.8 F 0.0
2:00 191.3 F 2.9
4:00 208.7 F 7.5
6:00 215.2 F 12.7
8:00 224.0 F 7.1
10:00 235.6 F 12.2
12:00 236.5 F 9.6
14:00 238.6 F 9.4
16:00 248.6 F 14.6
18:00 260.1 F 20.9
20:00 270.8 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.4
1:00 272.3 F 29.0
2:00 271.4 F 28.3
3:00 272.3 F 29.2
4:00 272.0 F 28.7

Min 271.2 F 28.3
Max 272.4 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.7 F 0.0
5:00 229.9 F 0.0
10:00 230.4 F 0.0
15:00 230.6 F 0.0
20:00 229.8 F 0.0
25:00 229.9 F 0.0
30:00 228.6 F 0.0
Drying Complete

FILLING: 1:43
HEATING: 20:46
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 00:59:54
CYCLE COMPLETE

01/17/2023 04:16 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____