

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.5uS / 0.3ppm  
CYCLE NUMBER 001404  
18:23 20/01/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:15  
PRESSURIZING 3:21  
STERILIZING 4:28  
136.0 C 220kPa 4:28  
136.4 C 224kPa 4:58  
136.3 C 224kPa 5:28  
136.2 C 222kPa 5:58  
136.4 C 224kPa 6:28  
136.4 C 223kPa 6:58  
136.0 C 222kPa 7:28  
136.4 C 224kPa 7:58  
136.4 C 224kPa 7:58  
VENTING 7:58  
AIR DRYING 8:53  
CYCLE COMPLETE 9:05  
-----

Digital Signature #  
D7341F3546AB4226  
-----