

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5491

Name: RITTER 1  
Sterilizer ID: V2178198  
04/09/2025 11:09 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 111.6 F 0.0  
2:00 113.0 F 1.2  
4:00 131.3 F 3.9  
6:00 159.3 F 7.1  
8:00 183.2 F 10.8  
10:00 212.4 F 3.6  
12:00 213.4 F 5.6  
14:00 224.5 F 9.5  
16:00 235.8 F 11.0  
18:00 238.9 F 10.0  
20:00 238.0 F 8.9  
22:00 246.9 F 13.4  
24:00 257.9 F 19.4  
26:00 268.2 F 25.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.5  
1:00 272.4 F 28.6  
2:00 272.2 F 29.0  
3:00 272.0 F 29.0  
4:00 271.9 F 28.5

Min 271.5 F 28.4  
Max 272.5 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 188.2 F 0.0  
5:00 209.3 F 0.0  
10:00 213.1 F 0.0  
15:00 217.9 F 0.0  
20:00 213.3 F 0.0  
25:00 219.1 F 0.0  
30:00 224.6 F 0.0  
Drying Complete

FILLING: 0:47  
HEATING: 27:55  
STERILIZING: 4:00  
VENTING: 3:43  
DRYING: 30:00  
TOTAL CYCLE: 01:06:26  
CYCLE COMPLETE

04/09/2025 12:15 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_