
Midmark M11 - v1.0.5
Total Cycles: 3932

Name: RITTER2
Sterilizer ID: V2313228
09/11/2024 04:45 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 114.1 F 0.0
2:00 115.6 F 0.0
4:00 120.5 F 0.4
6:00 150.0 F 3.3
8:00 177.3 F 6.6
10:00 209.2 F 10.2
12:00 223.7 F 7.9
14:00 227.1 F 9.5
16:00 236.7 F 10.4
18:00 238.7 F 11.4
20:00 236.7 F 9.7
22:00 249.6 F 15.7
24:00 260.0 F 21.9
26:00 270.2 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.4
1:00 271.6 F 29.9
2:00 272.7 F 30.0
3:00 271.5 F 29.4
4:00 271.4 F 29.9

Min 271.0 F 29.3
Max 272.7 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.0 F 0.0
5:00 217.6 F 0.0
10:00 234.2 F 0.0
15:00 231.5 F 0.0
20:00 222.1 F 0.0
25:00 220.1 F 0.0
30:00 222.5 F 0.0

Drying Complete

FILLING: 1:06
HEATING: 26:49
STERILIZING: 4:00
VENTING: 3:39
DRYING: 30:00
TOTAL CYCLE: 01:05:35
CYCLE COMPLETE

09/11/2024 05:50 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____