
Midmark M11 - v1.0.0
Total Cycles: 1681

Name: RITTER 1
Sterilizer ID: V2178198
02/28/2022 03:51 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 161.8 F 0.0
2:00 167.2 F 2.8
4:00 187.0 F 7.0
6:00 205.9 F 12.5
8:00 222.7 F 6.5
10:00 234.5 F 11.8
12:00 235.5 F 9.0
14:00 238.7 F 9.4
16:00 248.8 F 14.5
18:00 260.8 F 21.2
20:00 271.8 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.5 F 28.9
2:00 271.5 F 28.1
3:00 272.4 F 29.0
4:00 272.2 F 28.5

Min 271.4 F 28.1
Max 272.5 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.2 F 0.0
5:00 217.6 F 0.0
10:00 223.7 F 0.0
15:00 225.4 F 0.0
20:00 225.9 F 0.0
25:00 227.7 F 0.0
30:00 227.5 F 0.0
Drying Complete

FILLING: 1:29
HEATING: 20:35
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:59:13
CYCLE COMPLETE

02/28/2022 04:50 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____