
Midmark M11 - v1.0.0
Total Cycles: 4789

Name: RITTER 1
Sterilizer ID: V2178198
08/07/2024 12:41 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 110.7 F 0.0
2:00 119.8 F 0.1
4:00 134.5 F 0.3
6:00 168.7 F 0.7
8:00 189.6 F 2.6
10:00 203.4 F 5.9
12:00 215.3 F 9.3
14:00 226.0 F 13.1
16:00 229.0 F 7.7
18:00 239.6 F 12.1
20:00 237.3 F 9.3
22:00 240.2 F 10.1
24:00 246.9 F 13.4
26:00 257.9 F 19.3
28:00 267.5 F 25.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.1
1:00 272.1 F 28.7
2:00 272.5 F 28.8
3:00 272.1 F 28.5
4:00 271.5 F 28.1

Min 271.2 F 28.0
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.6 F 0.0
5:00 224.5 F 0.0
10:00 232.3 F 0.0
15:00 225.1 F 0.0
20:00 232.8 F 0.0
25:00 221.4 F 0.0
30:00 221.7 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 29:38
STERILIZING: 4:00
VENTING: 3:51
DRYING: 30:00
TOTAL CYCLE: 01:08:16
CYCLE COMPLETE

08/07/2024 01:49 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____