
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 002695
9:41 17/11/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:34
PRESSURIZING 2:39
STERILIZING 3:48
136.4 C 224kPa 3:48
136.2 C 223kPa 4:18
136.0 C 221kPa 4:48
136.2 C 223kPa 5:18
136.3 C 223kPa 5:48
136.2 C 221kPa 6:18
136.3 C 223kPa 6:48
136.4 C 223kPa 7:18
136.4 C 223kPa 7:18
VENTING 7:18
AIR DRYING 8:10
CYCLE COMPLETE 8:12

Digital Signature #
2DCCD314A1815894
