
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000651
8:48 13/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:03
PRESSURIZING 3:07
STERILIZING 4:17
136.3 C 230kPa 4:17
136.3 C 232kPa 4:47
136.4 C 232kPa 5:17
136.3 C 232kPa 5:47
136.3 C 233kPa 6:17
136.4 C 226kPa 6:47
136.4 C 231kPa 7:17
136.4 C 233kPa 7:47
136.4 C 233kPa 7:47
VENTING 7:48
AIR DRYING 8:43
CYCLE COMPLETE 8:49

Digital Signature #
6757A09915957343
