
Midmark M11 - v1.0.0
Total Cycles: 2459

Name: RITTER 1
Sterilizer ID: V2178198
10/19/2022 01:49 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 175.3 F 0.0
2:00 171.1 F 3.8
4:00 190.8 F 7.9
6:00 213.2 F 12.8
8:00 218.8 F 6.6
10:00 230.9 F 11.4
12:00 233.4 F 8.0
14:00 242.1 F 11.7
16:00 244.3 F 12.7
18:00 257.3 F 19.6
20:00 269.1 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.7
1:00 272.4 F 29.2
2:00 271.2 F 28.6
3:00 272.5 F 29.3
4:00 271.8 F 28.7

Min 271.2 F 28.5
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.3 F 0.0
5:00 235.7 F 0.0
10:00 236.1 F 0.0
15:00 231.1 F 0.0
20:00 230.4 F 0.0
25:00 230.0 F 0.0
30:00 227.4 F 0.0
Drying Complete

FILLING: 1:08
HEATING: 21:03
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:59:23
CYCLE COMPLETE

10/19/2022 02:48 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____