

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2989

Name: RITTER2  
Sterilizer ID: V2313228  
10/10/2023 01:38 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 164.6 F 0.0  
2:00 168.1 F 3.0  
4:00 195.0 F 7.4  
6:00 212.9 F 12.6  
8:00 225.8 F 7.2  
10:00 235.4 F 12.4  
12:00 235.8 F 9.7  
14:00 236.7 F 9.0  
16:00 248.3 F 14.9  
18:00 260.5 F 21.8  
20:00 271.8 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 29.3  
1:00 272.1 F 29.3  
2:00 272.1 F 29.6  
3:00 272.6 F 29.9  
4:00 271.7 F 29.5

Min 271.6 F 29.2  
Max 272.9 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 181.7 F 0.0  
5:00 236.7 F 0.0  
10:00 240.8 F 0.0  
15:00 231.0 F 0.0  
20:00 224.3 F 0.0  
25:00 237.1 F 0.0  
30:00 226.1 F 0.0  
Drying Complete

FILLING: 1:19  
HEATING: 21:15  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 00:59:48  
CYCLE COMPLETE

10/10/2023 02:37 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_