

-----  
Midmark M11 - v1.0.5  
Total Cycles: 421

Name: RITTER2  
Sterilizer ID: V2313228  
10/06/2021 11:44 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 124.8 F 0.0  
2:00 124.9 F 1.3  
4:00 139.8 F 3.6  
6:00 160.6 F 7.0  
8:00 187.7 F 11.2  
10:00 201.1 F 4.5  
12:00 221.7 F 8.9  
14:00 233.8 F 13.4  
16:00 234.8 F 10.1  
18:00 235.4 F 9.2  
20:00 246.0 F 13.6  
22:00 256.6 F 19.4  
24:00 266.2 F 25.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.0 F 28.9  
1:00 272.8 F 30.0  
2:00 272.4 F 29.7  
3:00 272.1 F 29.4  
4:00 272.1 F 29.7

Min 271.0 F 28.8  
Max 273.0 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 214.7 F 0.0  
5:00 210.3 F 0.0  
10:00 210.2 F 0.0  
15:00 206.3 F 0.0  
20:00 199.6 F 0.0  
25:00 197.0 F 0.0  
30:00 201.4 F 0.0  
Drying Complete

FILLING: 1:12  
HEATING: 25:45  
STERILIZING: 4:00  
VENTING: 3:20  
DRYING: 30:00  
TOTAL CYCLE: 01:04:17  
CYCLE COMPLETE

10/06/2021 01:03 PM  
ACCEPTED: X  
REJECTED:  
Operator: U01