
Midmark M11 - v1.0.5
Total Cycles: 3084

Name: RITTER2
Sterilizer ID: V2313228
11/03/2023 03:44 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 125.1 F 0.0
2:00 126.3 F 1.1
4:00 150.2 F 4.2
6:00 182.7 F 8.5
8:00 204.2 F 13.5
10:00 218.1 F 7.6
12:00 231.9 F 12.9
14:00 234.9 F 10.0
16:00 236.2 F 8.8
18:00 248.8 F 15.2
20:00 261.2 F 22.3
22:00 272.4 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 29.2
1:00 272.4 F 29.7
2:00 271.8 F 29.7
3:00 271.8 F 29.2
4:00 272.5 F 29.8

Min 271.5 F 29.0
Max 272.5 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.0 F 0.0
5:00 203.9 F 0.0
10:00 207.0 F 0.0
15:00 201.8 F 0.0
20:00 199.3 F 0.0
25:00 197.2 F 0.0
30:00 201.6 F 0.0
Drying Complete

FILLING: 1:00
HEATING: 22:17
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 01:00:37
CYCLE COMPLETE

11/03/2023 04:44 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____