
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 003088
9:57 07/02/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:26
PRESSURIZING 3:27
STERILIZING 4:37
136.5 C 220kPa 4:37
136.3 C 224kPa 5:07
136.4 C 220kPa 5:37
136.2 C 222kPa 6:07
136.4 C 221kPa 6:37
136.2 C 222kPa 7:07
136.4 C 223kPa 7:37
136.0 C 220kPa 8:07
136.0 C 220kPa 8:07
VENTING 8:07
AIR DRYING 8:59
CYCLE COMPLETE 9:18

Digital Signature #
FD23B660F5AECA42
