
Midmark M11 - v1.0.0
Total Cycles: 2937

Name: RITTER 1
Sterilizer ID: V2178198
02/23/2023 10:52 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 200.2 F 0.0
2:00 191.1 F 4.5
4:00 202.3 F 9.5
6:00 224.2 F 6.8
8:00 232.6 F 10.3
10:00 234.6 F 7.5
12:00 242.3 F 11.4
14:00 246.6 F 13.5
16:00 259.8 F 20.7
18:00 271.5 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.4 F 29.0
2:00 271.4 F 28.3
3:00 272.4 F 29.1
4:00 271.8 F 28.4

Min 271.4 F 28.2
Max 272.5 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.0 F 0.0
5:00 229.2 F 0.0
10:00 232.0 F 0.0
15:00 226.8 F 0.0
20:00 225.2 F 0.0
25:00 225.0 F 0.0
30:00 225.4 F 0.0
Drying Complete

FILLING: 1:48
HEATING: 18:39
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:57:40
CYCLE COMPLETE

02/23/2023 11:49 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____