
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002206
11:01 04/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:00
PRESSURIZING 2:04
STERILIZING 3:08
136.0 C 222kPa 3:08
136.4 C 224kPa 3:38
136.4 C 223kPa 4:08
136.1 C 223kPa 4:38
136.4 C 226kPa 5:08
136.4 C 226kPa 5:38
136.4 C 226kPa 6:08
136.4 C 225kPa 6:38
136.4 C 225kPa 6:38
VENTING 6:38
AIR DRYING 7:34
CYCLE COMPLETE 16:02

Digital Signature #
167A0A606B75DFBE
