
Midmark M11 - v1.0.0
Total Cycles: 3395

Name: RITTER 1
Sterilizer ID: V2178198
06/15/2023 10:56 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 193.0 F 0.0
2:00 193.5 F 0.9
4:00 206.1 F 6.1
6:00 221.4 F 11.3
8:00 228.4 F 7.4
10:00 241.4 F 13.1
12:00 240.9 F 11.3
14:00 239.4 F 10.2
16:00 253.6 F 17.2
18:00 265.6 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.4 F 28.7
2:00 272.1 F 28.9
3:00 272.0 F 28.4
4:00 272.5 F 29.1

Min 271.5 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.8 F 0.0
5:00 216.5 F 0.0
10:00 229.8 F 0.0
15:00 230.7 F 0.0
20:00 227.8 F 0.0
25:00 225.4 F 0.0
30:00 224.0 F 0.0
Drying Complete

FILLING: 1:36
HEATING: 19:39
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 00:58:36
CYCLE COMPLETE

06/15/2023 11:54 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____