
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001379
10:05 13/12/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:50
PRESSURIZING 3:52
STERILIZING 5:00
136.0 C 221kPa 5:00
136.3 C 224kPa 5:30
136.4 C 224kPa 6:00
136.4 C 225kPa 6:30
136.3 C 224kPa 7:00
136.2 C 221kPa 7:30
136.4 C 224kPa 8:00
136.4 C 224kPa 8:30
136.4 C 224kPa 8:30
VENTING 8:30
AIR DRYING 9:24
CYCLE COMPLETE 14:54

Digital Signature #
CC8B67280A7C334C
