
Midmark M11 - v1.0.5
Total Cycles: 3104

Name: RITTER2
Sterilizer ID: V2313228
11/09/2023 10:49 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 155.7 F 0.0
2:00 156.2 F 1.3
4:00 171.2 F 5.0
6:00 190.3 F 9.4
8:00 208.6 F 7.8
10:00 225.3 F 8.8
12:00 236.6 F 10.8
14:00 240.1 F 12.2
16:00 240.6 F 11.4
18:00 253.9 F 18.1
20:00 265.5 F 25.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.2
1:00 273.0 F 30.0
2:00 271.9 F 29.3
3:00 271.7 F 29.7
4:00 272.6 F 29.9

Min 271.2 F 29.1
Max 273.0 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.7 F 0.0
5:00 225.4 F 0.0
10:00 236.2 F 0.0
15:00 225.4 F 0.0
20:00 218.8 F 0.0
25:00 215.4 F 0.0
30:00 212.8 F 0.0
Drying Complete

FILLING: 1:16
HEATING: 21:48
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:00:36
CYCLE COMPLETE

11/09/2023 11:49 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____