
Midmark M11 - v1.0.0
Total Cycles: 5074

Name: RITTER 1
Sterilizer ID: V2178198
11/08/2024 01:32 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 119.2 F 0.0
2:00 121.8 F 0.1
4:00 149.6 F 0.5
6:00 177.5 F 1.5
8:00 206.1 F 5.7
10:00 219.1 F 9.7
12:00 229.1 F 10.6
14:00 229.7 F 8.5
16:00 241.5 F 13.3
18:00 239.2 F 10.8
20:00 236.3 F 9.0
22:00 250.7 F 15.8
24:00 262.4 F 22.5
26:00 272.7 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 272.5 F 29.2
2:00 271.7 F 28.4
3:00 272.1 F 29.2
4:00 272.3 F 29.0

Min 271.4 F 28.4
Max 272.5 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.4 F 0.0
5:00 200.1 F 0.0
10:00 201.5 F 0.0
15:00 200.4 F 0.0
20:00 193.1 F 0.0
25:00 194.4 F 0.0
30:00 196.5 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 26:20
STERILIZING: 4:00
VENTING: 3:43
DRYING: 30:00
TOTAL CYCLE: 01:04:50
CYCLE COMPLETE

11/08/2024 02:36 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____